



MRTU4/6 & ME5 TWO WAY RESET PINOUT


1. M12_4pin Female connector on cable.



2. M12_4pin Male connector on MRTU flowmeter body.



3. Terminal Blocks on ME5.



Pin number	Name	Description	Colour
1	RES	External Reset	Brown
2	N/C	No Connect	White
3	GND	Common Ground	Blue
4	P	Pulse Output	black